

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	127	
3 - 1	125	
4 - 1	131	1:2S
6 - 1	125	
7 - 1	127	
8 - 1	128	
10 - 1	127	
11 - 1	127	
13 - 1	126	
14 - 1	128	
15 - 1	128	
15 - 1	128	
16 - 1	128	
20 - 1	128	
21 - 1	130	1:2S
22 - 1	127	
23 - 1	124	
24 - 1	130	1:2S
28 - 1	127	
29 - 1	126	
30 - 1	128	

Westgard	تفسیر	Copper (Cu)	آزمایش	اتوانالایزر BS-800	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	126.71	SD	1.59	Exp. Date	1402/9/20	Lot. No	hsn.35
Next Month	Mean	127.38	SD	1.69	CV	1.32			

