

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	0.45
SD	0.02
CV	5.05
Day-Run	Value
1 - 1	0.46
3 - 1	0.4
4 - 1	0.44
6 - 1	0.5
7 - 1	0.46
8 - 1	0.49
10 - 1	0.42
11 - 1	0.43
13 - 1	0.46
14 - 1	0.43
15 - 1	0.46
16 - 1	0.48
20 - 1	0.46
21 - 1	0.48
22 - 1	0.43
23 - 1	0.49
24 - 1	0.46
24 - 1	0.46
28 - 1	0.43
29 - 1	0.4
30 - 1	0.47

High - CV : 3.97	
Mean	2.21
SD	0.09
CV	3.97
Day-Run	Value
1 - 1	2.2
3 - 1	2.03
4 - 1	2.27
6 - 1	2.07
7 - 1	2.18
8 - 1	2.28
10 - 1	2.07
11 - 1	2.3
13 - 1	2.25
14 - 1	2.1
15 - 1	2.24
16 - 1	2.26
20 - 1	2.3
21 - 1	2.26
22 - 1	2.17
23 - 1	2.31
24 - 1	2.24
28 - 1	2.29
29 - 1	2.11
30 - 1	2.2

Westgard	تفسیر	D.Bili	آزمایش	اتوآنالایزر BS-800	دستگاه	8	ماه	1402	سال
Normal Control	biorex	Mean	0.46	SD	0.02	Exp. Date	1403/1/17	Lot. No	035

