

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	129.9	
2 - 1	131.8	
3 - 1	131.2	
4 - 1	125.7	
6 - 1	127	
7 - 1	135.2	
8 - 1	125.3	
9 - 1	127.5	
10 - 1	125	
11 - 1	131.2	
20 - 1	135	
21 - 1	130.9	
22 - 1	135.8	
23 - 1	131.2	
24 - 1	129.5	
25 - 1	128.5	
27 - 1	135.9	
28 - 1	134.2	
29 - 1	136	
30 - 1	128.9	4:1S

Westgard	تفسیر	Ferritin	آزمایش	SIEMENS Centaur CP	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	129.87	SD	3.16	Exp. Date	1403/2/10	Lot. No	365-02-028
Next Month	Mean	131.09	SD	3.74	CV	2.86			

