

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	7.25	
2 - 1	7.5	
3 - 1	7.32	
4 - 1	7.25	
6 - 1	7.42	
7 - 1	7.46	
8 - 1	7.48	
9 - 1	7.56	1:3S
10 - 1	7.18	1:2S
11 - 1	7.29	
20 - 1	7.56	
21 - 1	7.48	
22 - 1	7.58	
23 - 1	7.46	
24 - 1	7.45	
25 - 1	7.33	
27 - 1	7.53	
28 - 1	7.6	1:2S
29 - 1	7.45	
30 - 1	7.35	

Westgard	تفسیر	Free T3	آزمایش	SIEMENS Centaur CP	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	7.4	SD	0.09	Exp. Date	1401/12/25	Lot. No	36-02-026
Next Month	Mean	7.43	SD	0.12	CV	1.64			

