

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	67.1	
3 - 1	67	
4 - 1	67	
6 - 1	66.1	
7 - 1	70.1	
8 - 1	70.6	1:2S
10 - 1	67.5	
11 - 1	69.2	
13 - 1	68.8	
14 - 1	69.3	
15 - 1	70.6	1:2S 4:1S
16 - 1	66	
20 - 1	67.6	
21 - 1	70.1	
22 - 1	69.9	
23 - 1	68.6	
24 - 1	69.2	
28 - 1	68.5	
29 - 1	71.1	1:2S
30 - 1	66	

Westgard	تفسیر	Gamma GT	آزمایش	اتوآنالایزر BS-800	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	68	SD	1.77	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	68.52	SD	1.64	CV	2.39			

