

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	137.6	
3 - 1	135.8	
4 - 1	139.4	
6 - 1	133	
7 - 1	135.7	
8 - 1	140	
10 - 1	138.4	
11 - 1	140	
13 - 1	137.8	
14 - 1	140	
15 - 1	139.8	
16 - 1	139.3	
20 - 1	136	
21 - 1	133.5	
22 - 1	134.8	
23 - 1	133.2	
24 - 1	137.1	
28 - 1	142	1:2S
29 - 1	135.5	
30 - 1	135.1	

Westgard	تفسیر	LDL	آزمایش	اتوآنالایزر BS-800	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	136.45	SD	2.63	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	137.2	SD	2.62	CV	1.91			

