

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	36	
2 - 1	36	
3 - 1	36.3	
4 - 1	36.2	
5 - 1	36	
6 - 1	36.3	
7 - 1	36.1	
8 - 1	36	
9 - 1	36.3	
10 - 1	36.5	
11 - 1	35.9	
12 - 1	36.2	
13 - 1	36.4	
14 - 1	36	
15 - 1	36.6	
16 - 1	36.2	
17 - 1	36.4	
18 - 1	36.1	
19 - 1	36.4	
20 - 1	36	
21 - 1	36.5	
22 - 1	36.5	
23 - 1	36.6	
24 - 1	36.8	
25 - 1	36	
28 - 1	36.2	
29 - 1	36.5	
30 - 1	36.5	

Westgard	تفسیر	MCHC	آزمایش	سل کانتیر Siemens Advia	دستگاه	8	ماه	1402	سال
Control	R&D	Mean	36.25	SD	0.57	Exp. Date	1402/8/25	Lot. No	
Next Month	Mean	36.27	SD	0.24	CV	0.66			

