

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	41.6	
2 - 1	40.65	
8 - 1	40	
8 - 2	40.39	
8 - 3	40.12	
10 - 1	39.9	
12 - 1	40.09	
15 - 1	40.1	
16 - 1	40.54	
18 - 1	40.5	
20 - 1	40.3	
21 - 1	41.3	
23 - 1	41.08	
27 - 1	41.01	
29 - 1	40.3	
30 - 1	40.09	

Westgard	تفسیر	PTH intact	آزمایش	Euroimmun analyser 1	دستگاه	8	ماه	1402	سال
Control	Monobind	Mean	40.42	SD	0.64	Exp. Date	1402/12/24	Lot. No	
Next Month	Mean	40.5	SD	0.5	CV	1.24			

