

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.12	
2 - 1	4.15	
4 - 1	4.3	
6 - 1	4.2	
7 - 1	4	
8 - 1	4.12	
7 - 2	4.2	
8 - 2	4.3	
9 - 1	4.1	
10 - 1	4.01	
13 - 1	4.3	
14 - 1	4.2	
15 - 1	4.3	
16 - 1	4	
17 - 1	4.2	
18 - 1	4.3	
20 - 1	4.2	
21 - 1	4	
23 - 1	4.3	
24 - 1	4.1	
25 - 1	4	
27 - 1	4.3	
28 - 1	4.2	
29 - 1	4	
30 - 1	4.17	

Westgard	تفسیر	RBC	آزمایش	سل کانتور Siemens Advia	دستگاه	8	ماه	1402	سال
Control	R&D	Mean	4.13	SD	0.11	Exp. Date	1402/8/25	Lot. No	
Next Month	Mean	4.16	SD	0.11	CV	2.72			

