

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.6	
2 - 1	2.37	
3 - 1	2.63	
4 - 1	2.55	
6 - 1	2.68	
7 - 1	2.5	
8 - 1	2.76	
9 - 1	2.65	
10 - 1	2.6	
11 - 1	2.55	
20 - 1	2.46	
21 - 1	2.29	
22 - 1	2.18	1:2S
23 - 1	2.37	
24 - 1	2.22	
25 - 1	2.58	
27 - 1	2.42	
28 - 1	2.2	
29 - 1	2.25	
30 - 1	2.62	

Westgard	تفسیر	T3	آزمایش	SIEMENS Centaur CP	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	2.51	SD	0.16	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	2.47	SD	0.18	CV	7.13			

