

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 1 - 1 | 323 | |
| 3 - 1 | 301 | |
| 4 - 1 | 300.7 | |
| 6 - 1 | 303.4 | |
| 7 - 1 | 303.6 | |
| 8 - 1 | 303 | |
| 10 - 1 | 294 | |
| 11 - 1 | 297.4 | |
| 13 - 1 | 300 | |
| 14 - 1 | 297.2 | |
| 15 - 1 | 305 | 10X |
| 16 - 1 | 304.4 | |
| 20 - 1 | 305.4 | |
| 21 - 1 | 306.1 | |
| 23 - 1 | 304.9 | |
| 24 - 1 | 306.6 | |
| 28 - 1 | 300 | |
| 29 - 1 | 300 | |
| 30 - 1 | 295.9 | |

| | | | | | | | | | |
|------------|--------|--------|--------|--------------------|--------|-----------|-----------|---------|-----|
| Westgard | تفسیر | TIBC | آزمایش | BS-800 اتوآنالایزر | دستگاه | 8 | ماه | 1402 | سال |
| Control | biorex | Mean | 304 | SD | 9.3 | Exp. Date | 1403/1/17 | Lot. No | 035 |
| Next Month | Mean | 302.72 | SD | 6.1 | CV | 2.01 | | | |

