

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	1.92	
2 - 1	1.89	
2 - 2	2.1	
2 - 3	1.92	
2 - 4	1.96	
5 - 1	2.21	
8 - 1	1.98	
9 - 1	1.99	
10 - 1	2.01	
11 - 1	1.8	
11 - 2	2.62	
13 - 1	1.69	
16 - 1	1.89	
17 - 1	2.01	
19 - 1	1.87	
22 - 1	2.31	
25 - 1	2.24	
26 - 1	2.16	
27 - 1	2.16	
29 - 1	1.97	
30 - 1	2.03	

Westgard	تفسیر	TSH	آزمایش	Euroimmun analyser 1	دستگاه	8	ماه	1401	سال
Control		Mean	2.05	SD	0.4	Exp. Date	1402/8/30	Lot. No	
Next Month		Mean	2.03	SD	0.2	CV	9.9		

