

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.84	
Mean	4.23	Mean	8.67
SD	0.06	SD	0.16
CV	1.35	CV	1.84
Day-Run	Value	Day-Run	Value
1 - 1	4.2	1 - 1	8.8
3 - 1	4.3	3 - 1	8.6
4 - 1	4.2	4 - 1	9
6 - 1	4.2	6 - 1	8.7
7 - 1	4.2	7 - 1	8.7
8 - 1	4.3	8 - 1	8.6
10 - 1	4.2	10 - 1	8.6
11 - 1	4.3	11 - 1	8.5
13 - 1	4.3	13 - 1	8.6
14 - 1	4.3	14 - 1	8.6
15 - 1	4.2	15 - 1	8.7
16 - 1	4.3	16 - 1	8.8
20 - 1	4.2	20 - 1	8.5
21 - 1	4.1	21 - 1	8.5
22 - 1	4.2	22 - 1	8.6
23 - 1	4.2	23 - 1	8.9
24 - 1	4.2	24 - 1	8.5
28 - 1	4.3	28 - 1	8.9
29 - 1	4.2	29 - 1	8.4
30 - 1	4.2	30 - 1	8.8

Westgard	تفسیر	Uric Acid	آزمایش	اتوانالایزر BS-800	دستگاه	8	ماه	1402	سال
Normal Control	biorex	Mean	4.2	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

