

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	29.6	
2 - 2	28.9	
3 - 1	29.1	
4 - 1	29.6	
6 - 1	28.9	
7 - 1	29	
8 - 1	29.3	
9 - 1	29.2	
10 - 1	29	
11 - 1	29.2	
14 - 1	29.4	
14 - 2	28.9	
15 - 1	29.2	
16 - 1	29.1	
17 - 1	29.7	
18 - 1	29.2	
20 - 1	29.1	
21 - 1	29.8	
22 - 1	28.9	
23 - 1	28.7	
24 - 1	29.5	
25 - 1	28.8	
27 - 1	29.4	
28 - 1	28.9	
29 - 1	29	
30 - 1	29.4	

Westgard	تفسیر	V D3	آزمایش	HPLC Agilent 1100 -1	دستگاه	8	ماه	1402	سال
Control	آزمایشگاه	Mean	29.22	SD	0.41	Exp. Date	1402/8/30	Lot. No	
Next Month	Mean	29.18	SD	0.29	CV	1.01			

