

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 1 - 1   | 6.32  |             |
| 2 - 1   | 6.4   |             |
| 3 - 1   | 6.2   |             |
| 4 - 1   | 6.3   |             |
| 6 - 1   | 6.2   |             |
| 7 - 1   | 6.4   |             |
| 8 - 1   | 6.32  |             |
| 9 - 1   | 6.2   |             |
| 10 - 1  | 6.3   |             |
| 11 - 1  | 6.4   |             |
| 13 - 1  | 6.3   |             |
| 15 - 1  | 6.2   |             |
| 16 - 1  | 6.3   |             |
| 17 - 1  | 6.3   |             |
| 18 - 1  | 6.4   |             |
| 18 - 2  | 6.32  |             |
| 19 - 1  | 6.32  |             |
| 20 - 1  | 6.2   |             |
| 21 - 1  | 6.4   |             |
| 22 - 1  | 6.32  |             |
| 24 - 1  | 6.3   |             |
| 25 - 1  | 6.1   |             |
| 27 - 1  | 6.3   |             |
| 28 - 1  | 6.32  |             |
| 29 - 1  | 6.4   |             |
| 30 - 1  | 6.5   |             |

|            |       |      |        |                         |        |           |           |         |     |
|------------|-------|------|--------|-------------------------|--------|-----------|-----------|---------|-----|
| Westgard   | تفسیر | WBC  | آزمایش | سل کانتیر Siemens Advia | دستگاه | 8         | ماه       | 1402    | سال |
| Control    | R&D   | Mean | 6.32   | SD                      | 0.19   | Exp. Date | 1402/8/25 | Lot. No |     |
| Next Month | Mean  | 6.31 | SD     | 0.09                    | CV     | 1.38      |           |         |     |

