

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.84	
Mean	4.17	Mean	5.08
SD	0.08	SD	0.09
CV	1.86	CV	1.84
Day-Run	Value	Day-Run	Value
1 - 1	4.2	1 - 1	5.1
2 - 1	4.2	2 - 1	5.1
5 - 1	4.3	5 - 1	5.1
6 - 1	4.2	6 - 1	5
7 - 1	4.2	7 - 1	5.1
8 - 1	4.1	8 - 1	5
9 - 1	4.1	9 - 1	5
11 - 1	4.1	11 - 1	5.2
12 - 1	4.2	12 - 1	5.1
13 - 1	4.1	13 - 1	5.1
14 - 1	4.1	14 - 1	5.1
15 - 1	4.1	15 - 1	5.2
16 - 1	4.2	16 - 1	5
18 - 1	4.1	18 - 1	5
19 - 1	4.1	19 - 1	5.1
20 - 1	4	20 - 1	5.1
21 - 1	4.1	21 - 1	4.9
22 - 1	4.2	22 - 1	4.9
23 - 1	4.2	23 - 1	5
25 - 1	4.3	25 - 1	5.2
27 - 1	4.2	27 - 1	5.2
28 - 1	4.3	28 - 1	5.2
29 - 1	4.2	29 - 1	5.2

Westgard	تفسیر	Albomin	آزمایش	BS-800 اتوآنالایزر	دستگاه	9	ماه	1402	سال
Normal Control	biorex	Mean	4.2	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

