

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	24.7	
6 - 1	24.4	
7 - 1	27.4	
8 - 1	22.3	
9 - 1	22	
11 - 1	25.2	
12 - 1	24.3	
13 - 1	25	
14 - 1	23.8	
15 - 1	24.7	
16 - 1	23.5	
18 - 1	25	
19 - 1	23.2	
20 - 1	26.1	
21 - 1	23.1	
22 - 1	23.7	
25 - 1	24.9	
27 - 1	24	
28 - 1	26.2	
29 - 1	24	
30 - 1	25.4	

Westgard	تفسیر	BHCG	آزمایش	2000 ایمولایت XPI-	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	24.54	SD	1.47	Exp. Date	1403/10/1	Lot. No	
Next Month	Mean	24.42	SD	1.29	CV	5.28			

