

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.74	
Mean	9.37	Mean	13.52
SD	0.1	SD	0.23
CV	1.06	CV	1.74
Day-Run	Value	Day-Run	Value
1 - 1	9.3	1 - 1	13.5
1 - 1	9.3	2 - 1	13.6
2 - 1	9.5	5 - 1	13.3
6 - 1	9.3	11 - 1	13.8
11 - 1	9.4	12 - 1	13.4
12 - 1	9.3	13 - 1	13.5
13 - 1	9.4	14 - 1	13.6
14 - 1	9.3	15 - 1	13.7
15 - 1	9.3	16 - 1	13.6
16 - 1	9.3	18 - 1	13.7
18 - 1	9.4	19 - 1	13.7
19 - 1	9.5	20 - 1	13.8
20 - 1	9.4	21 - 1	13.6
21 - 1	9.5	22 - 1	13.6
22 - 1	9.2	23 - 1	13.8
23 - 1	9.4	25 - 1	13
25 - 1	9.3	27 - 1	13.3
27 - 1	9.3	28 - 1	13.4
28 - 1	9.3	28 - 2	13.4
29 - 1	9.6	29 - 1	13

Westgard	تفسیر	Calcium	آزمایش	BS-800 اتوآنالایزر	دستگاه	9	ماه	1402	سال
Normal Control	biorex	Mean	9.4	SD	0.12	Exp. Date	1403/1/17	Lot. No	035

