

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	73.33
SD	1.08
CV	1.48
Day-Run	Value
1 - 1	74
2 - 1	74
5 - 1	73
6 - 1	74
7 - 1	72
8 - 1	72
9 - 1	72
11 - 1	73
12 - 1	74
13 - 1	75
14 - 1	73
15 - 1	72
16 - 1	72
18 - 1	73
19 - 1	72
20 - 1	73
21 - 1	74
22 - 1	74.5
23 - 1	75
25 - 1	74
27 - 1	72
28 - 1	74
29 - 1	75

High - CV : 1.61	
Mean	229.57
SD	3.69
CV	1.61
Day-Run	Value
1 - 1	230
1 - 2	230
5 - 1	223
6 - 1	220
7 - 1	226
8 - 1	228
9 - 1	235
11 - 1	230
12 - 1	230
13 - 1	231
14 - 1	226
15 - 1	232
16 - 1	229
18 - 1	231
19 - 1	230
20 - 1	232
21 - 1	234
22 - 1	230
23 - 1	232
25 - 1	227
27 - 1	229
28 - 1	237
29 - 1	228

Westgard	تفسیر	FBS	آزمایش	اتوآنالایزر BS-800	دستگاه	9	ماه	1402	سال
Normal Control	biorex	Mean	73	SD	0.86	Exp. Date	1403/1/17	Lot. No	035

