

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
18 - 1	141.3	
19 - 1	138.1	
20 - 1	140.1	
21 - 1	145.7	
22 - 1	142	
23 - 1	141.7	
25 - 1	143.5	
27 - 1	138.5	
28 - 1	150.2	
29 - 1	153.9	
30 - 1	158.2	1:2S

Westgard	تفسیر	Ferritin	آزمایش	SIEMENS Centaur CP	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	147.44	SD	4.92	Exp. Date	1403/2/10	Lot. No	365-02-028
Next Month	Mean	144.84	SD	6.56	CV	4.53			

