

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	766	
2 - 1	774.1	
4 - 1	780	
5 - 1	770.2	
8 - 1	762.3	
9 - 1	760	
11 - 1	774.1	
12 - 1	783.1	
13 - 1	755.2	
14 - 1	770.5	
15 - 1	750.2	
16 - 1	770	
18 - 1	775.2	
19 - 1	761.3	
20 - 1	772.2	
21 - 1	778.3	1:2S
22 - 1	762.5	
23 - 1	761.5	
25 - 1	759.3	
27 - 1	767.2	
28 - 1	770	
29 - 1	780.3	
30 - 1	761.3	

Westgard	تفسیر	Ig G	آزمایش	BS-480 اتوآنالایزر	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	763.08	SD	10.08	Exp. Date	1402/11/7	Lot. No	641-01-030/2
Next Month	Mean	768.03	SD	8.56	CV	1.11			

