

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	27.9	1:2S
4 - 1	29.2	
5 - 1	30.1	
6 - 1	31.3	
7 - 1	32.5	
8 - 1	30.3	
9 - 1	29.9	
11 - 1	33.1	
12 - 1	32.1	
14 - 1	32.3	
15 - 1	30.3	
16 - 1	30.8	
18 - 1	29.8	
19 - 1	29.4	
20 - 1	34	
21 - 1	29.4	
22 - 1	32.1	
25 - 1	35.9	1:2S
27 - 1	34.4	
28 - 1	31	
30 - 1	31.9	

Westgard	تفسیر	LH	آزمایش	2000 ایمولایت XPI-	دستگاه	9	ماه	1402	سال
Control	Seronorm	Mean	32.3	SD	1.55	Exp. Date	1403/5/31	Lot. No	2011909
Next Month	Mean	31.32	SD	1.96	CV	6.25			

