

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
4 - 1	43.6	
5 - 1	43.8	
8 - 1	42.8	
9 - 1	43.6	
11 - 1	44.1	
12 - 1	44.3	
13 - 1	43.7	
14 - 1	43.7	
15 - 1	42.7	
16 - 1	43.7	
18 - 1	44.1	
19 - 1	42.9	
20 - 1	43.6	
21 - 1	43.8	
22 - 1	42.89	
23 - 1	44.3	
25 - 1	42.8	
27 - 1	43.2	1:2S
28 - 1	44.8	1:2S
29 - 1	42.6	
30 - 1	43.7	

Westgard	تفسیر	Lactate(Lactic Acid)	آزمایش	BS-480 اتوآنالایزر	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	43.54	SD	0.55	Exp. Date	1402/12/14	Lot. No	
Next Month	Mean	43.56	SD	0.6	CV	1.38			

