

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	17.2	
4 - 1	18.7	
5 - 1	18.5	
6 - 1	18.2	
7 - 1	19.3	
8 - 1	16.6	1:2S
9 - 1	17.2	
11 - 1	17.5	
12 - 1	18.8	
13 - 1	17.6	
14 - 1	17.5	
15 - 1	17.9	
16 - 1	16.4	1:2S
18 - 1	19.3	
19 - 1	18.1	
20 - 1	18.7	
21 - 1	17.2	
22 - 1	17.4	
25 - 1	19.8	1:2S
27 - 1	18.9	
29 - 1	19.2	
30 - 1	19.7	

Westgard	تفسیر	Prolactin	آزمایش	2000 ایمولایت XPI-	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	18.16	SD	0.77	Exp. Date	1403/2/10	Lot. No	365-02-028
Next Month	Mean	18.17	SD	0.99	CV	5.44			

