

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.83	
4 - 1	3.18	
5 - 1	2.94	
6 - 1	2.83	
7 - 1	2.9	
8 - 1	2.69	
9 - 1	2.96	
11 - 1	2.91	
12 - 1	2.98	
13 - 1	3.24	
14 - 1	3.1	
15 - 1	3.09	
16 - 1	3.07	
18 - 1	3.17	
19 - 1	3.23	
20 - 1	2.98	
21 - 1	2.91	
22 - 1	2.88	
25 - 1	3.17	
27 - 1	3.2	
29 - 1	3.37	
30 - 1	2.7	

Westgard	تفسیر	PSA	آزمایش	2000 ایمولایت XPI-	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	3.09	SD	0.2	Exp. Date	1403/10/1	Lot. No	365-02-028
Next Month	Mean	3.02	SD	0.18	CV	6			

