

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.73	
2 - 1	4.8	
4 - 1	4.62	
5 - 1	4.72	
8 - 1	4.8	
8 - 2	4.7	
11 - 1	4.78	
12 - 1	4.67	
13 - 1	4.68	
14 - 1	4.8	
15 - 1	4.79	
16 - 1	4.81	
18 - 1	4.76	
19 - 1	4.83	
20 - 1	4.6	
21 - 1	4.72	
22 - 1	4.8	
23 - 1	4.75	
25 - 1	4.83	

Westgard	تفسیر	Potassium (K)	آزمایش	AUDICOM	دستگاه	9	ماه	1402	سال
Control	زیست شیمی	Mean	4.71	SD	0.06	Exp. Date	1404/1/14	Lot. No	210202-17
Next Month	Mean	4.75	SD	0.07	CV	1.44			

