

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	111.3	
2 - 1	110.5	
4 - 1	109.5	
5 - 1	104.5	
8 - 1	106.2	
9 - 1	107.5	
11 - 1	109.1	
12 - 1	105.3	
13 - 1	104.7	
14 - 1	106.2	
15 - 1	111.6	
16 - 1	104.3	
18 - 1	107	
19 - 1	103	
20 - 1	104.2	
21 - 1	106.7	
22 - 1	108.3	
23 - 1	107.8	
25 - 1	105.8	
27 - 1	108.3	
28 - 1	106.9	
29 - 1	108	
30 - 1	105.8	

Westgard	تفسیر	RF	آزمایش	BS-480 اتوآنالایزر	دستگاه	9	ماه	1402	سال
Control	aptec	Mean	107.61	SD	2.53	Exp. Date	1403/8/19	Lot. No	06095
Next Month	Mean	107.07	SD	2.31	CV	2.16			

