

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	3.29	1:3S
2 - 1	3.65	
4 - 1	4.13	
5 - 1	3.35	1:2S
6 - 1	3.82	
7 - 1	4.06	
8 - 1	3.95	
9 - 1	3.72	
11 - 1	4.17	
12 - 1	3.8	
13 - 1	3.62	
14 - 1	3.25	1:2S
15 - 1	3.45	
16 - 1	3.55	
18 - 1	3.8	
19 - 1	3.9	
20 - 1	3.75	
21 - 1	4.02	
22 - 1	3.7	
23 - 1	3.45	
25 - 1	3.25	1:2S
27 - 1	3.65	1:3S
28 - 1	3.55	
29 - 1	3.98	
30 - 1	3.8	

Westgard	تفسیر	TSH	آزمایش	SIEMENS Centaur CP	دستگاه	9	ماه	1402	سال
Control	biorex	Mean	3.81	SD	0.2	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	3.71	SD	0.27	CV	7.27			

