

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	539	
4 - 1	576	
5 - 1	589	
6 - 1	553	
7 - 1	473	
8 - 1	485	
9 - 1	557	
11 - 1	524	
13 - 1	476	
14 - 1	410	1:2S
18 - 1	522	
19 - 1	438	1:2S
20 - 1	575	
21 - 1	429	1:2S
22 - 1	452	
25 - 1	512	
27 - 1	446	1:2S
29 - 1	450	
30 - 1	525	

Westgard	تفسیر	Testosterone	آزمایش	2000 ایمولایت XPI-	دستگاه	9	ماه	1402	سال
Control	Labex	Mean	536	SD	43.55	Exp. Date	1403/10/1	Lot. No	36502028
Next Month	Mean	501.63	SD	55.23	CV	11.01			

