

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 3.36	
Mean	4.17	Mean	8.84
SD	0.07	SD	0.3
CV	1.61	CV	3.36
Day-Run	Value	Day-Run	Value
1 - 1	4.2	1 - 1	8.5
2 - 1	4.2	2 - 1	8.8
11 - 1	4.2	11 - 1	8.9
12 - 1	4.2	12 - 1	9
13 - 1	4.2	13 - 1	8.5
14 - 1	4.2	14 - 1	9.2
15 - 1	4.1	15 - 1	9.2
16 - 1	4.2	16 - 1	8.4
18 - 1	4.1	18 - 1	9
19 - 1	4	19 - 1	8.8
20 - 1	4.2	20 - 1	8.9
21 - 1	4.1	21 - 1	9
21 - 2	4.1	21 - 2	9
22 - 1	4.2	22 - 1	8.9
23 - 1	4.1	23 - 1	9
25 - 1	4.2	25 - 1	8
27 - 1	4.3	27 - 1	9
28 - 1	4.2	28 - 1	8.8
29 - 1	4.2	29 - 1	9

Westgard	تفسیر	Uric Acid	آزمایش	اتوآنالایزر BS-800	دستگاه	9	ماه	1402	سال
Normal Control	biorex	Mean	4.2	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

