

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.57	
Mean	33.44	Mean	131.1
SD	0.9	SD	2.05
CV	2.71	CV	1.57
Day-Run	Value	Day-Run	Value
1 - 1	33.4	1 - 1	130.5
2 - 1	34.2	2 - 1	128.5
5 - 1	32.8	5 - 1	132
6 - 1	33.9	6 - 1	129
9 - 1	32	9 - 1	133
11 - 1	33	11 - 1	134
12 - 1	35.2	12 - 1	126.7
13 - 1	32.8	13 - 1	133.8
14 - 1	35	14 - 1	129.2
15 - 1	33	15 - 1	133.6
16 - 1	33.1	16 - 1	131.8
18 - 1	33.1	18 - 1	133
19 - 1	33.8	19 - 1	130
20 - 1	33.5	20 - 1	132
21 - 1	32.5	21 - 1	133
22 - 1	33	22 - 1	132
23 - 1	32.4	23 - 1	130
25 - 1	32.6	25 - 1	128
27 - 1	34	27 - 1	130
28 - 1	33.8	28 - 1	131
29 - 1	35.2	29 - 1	132.1

Westgard	تفسیر	Urea	آزمایش	اتوآنالایزر BS-800	دستگاه	9	ماه	1402	سال
Normal Control	biorex	Mean	34.46	SD	1.12	Exp. Date	1403/1/17	Lot. No	035

