

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	174.64
SD	3
CV	1.72
Day-Run	Value
1 - 1	177
2 - 1	178
3 - 1	172
4 - 1	170.5
7 - 1	181
8 - 1	177
9 - 1	170
10 - 1	175.9
11 - 1	172.1
12 - 1	175
14 - 1	176.6
15 - 1	173.6
16 - 1	171.7
17 - 1	174.1
18 - 1	172
21 - 1	173
23 - 1	178
24 - 1	170
25 - 1	177
28 - 1	174.2
29 - 1	175.8
30 - 1	177.6

High - CV : 1.52	
Mean	476.13
SD	7.25
CV	1.52
Day-Run	Value
1 - 1	460
2 - 1	470
3 - 1	482
4 - 1	480.1
7 - 1	480
8 - 1	466
9 - 1	485
10 - 1	470
11 - 1	476
12 - 1	473
14 - 1	469.7
15 - 1	480
16 - 1	482
17 - 1	470
18 - 1	470
21 - 1	480
23 - 1	476
24 - 1	482
25 - 1	486
28 - 1	473.3
29 - 1	470
30 - 1	490

Westgard	تفسیر	Ikalan Phosphatase(ALP)	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	173.09	SD	3.41	Exp. Date	1403/1/17	Lot. No	035

