

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	173.09
SD	3.41
CV	1.97
Day-Run	Value
2 - 1	171
3 - 1	177
4 - 1	171
5 - 1	172.6
6 - 1	175
7 - 1	173
9 - 1	173.6
10 - 1	169.6
11 - 1	170
12 - 1	180
13 - 1	170
14 - 1	170
16 - 1	170
17 - 1	173.6
18 - 1	170
19 - 1	168
20 - 1	180
21 - 1	171
23 - 1	174.2
24 - 1	173
25 - 1	179.2
26 - 1	175.6
27 - 1	171
28 - 1	172
30 - 1	176.8

High - CV : 1.42	
Mean	472.32
SD	6.71
CV	1.42
Day-Run	Value
2 - 1	459
3 - 1	472
4 - 1	469
5 - 1	474.6
6 - 1	473
7 - 1	467
9 - 1	480
10 - 1	473
11 - 1	477
12 - 1	475
13 - 1	480
14 - 1	470
16 - 1	470
17 - 1	474
18 - 1	463.5
19 - 1	488
20 - 1	463.5
21 - 1	483
23 - 1	481.1
24 - 1	467.3
25 - 1	472.1
26 - 1	466.6
27 - 1	474
28 - 1	468
30 - 1	467.4

Westgard	تفسیر	Ikalan Phosphatase(ALP)	آزمایش	اتوآنالایزر BS-800	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	174.4	SD	3.74	Exp. Date	1403/1/17	Lot. No	035

