

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	38.8
SD	0.79
CV	2.03
Day-Run	Value
1 - 1	39.1
2 - 1	39
3 - 1	37.8
4 - 1	38.7
7 - 1	40.1
8 - 1	38
9 - 1	39
10 - 1	39.1
11 - 1	38.7
12 - 1	38
14 - 1	38.7
15 - 1	38.9
16 - 1	38
17 - 1	40
18 - 1	37
21 - 1	39.9
23 - 1	38.3
24 - 1	39.7
25 - 1	39.4
28 - 1	38.4
29 - 1	38.4
30 - 1	39.5

High - CV : 1.11	
Mean	147.66
SD	1.63
CV	1.11
Day-Run	Value
1 - 1	150.9
2 - 1	151
3 - 1	147.8
4 - 1	146.2
7 - 1	145
8 - 1	149
9 - 1	147
10 - 1	147
11 - 1	148
12 - 1	149
14 - 1	145
15 - 1	147
16 - 1	149
17 - 1	149
18 - 1	148
21 - 1	147.4
23 - 1	147
24 - 1	149.2
25 - 1	146.8
28 - 1	147.2
29 - 1	145.9
30 - 1	146.1

Westgard	تفسیر	SGOT (AST)	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	38.39	SD	0.77	Exp. Date	1403/1/17	Lot. No	035

