

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	91	
2 - 1	91.8	
3 - 1	90.2	
4 - 1	90.8	
7 - 1	90.6	
8 - 1	92	
9 - 1	90.6	
10 - 1	91.4	
12 - 1	92.3	
14 - 1	91	
15 - 1	91.3	
16 - 1	92	
17 - 1	91	
18 - 1	92	
21 - 1	90	
23 - 1	91.8	
24 - 1	92.4	
25 - 1	90	
28 - 1	92.2	
29 - 1	89.9	
30 - 1	90.4	

Westgard	تفسیر	C3	آزمایش	BS-480 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	91.1	SD	0.69	Exp. Date	1402/11/7	Lot. No	641-01-030/2
Next Month	Mean	91.18	SD	0.82	CV	0.9			

