

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	91.9	
3 - 1	90.9	
4 - 1	90.8	
5 - 1	91.8	
6 - 1	91.5	
7 - 1	90.5	
9 - 1	90.3	
10 - 1	91.6	
11 - 1	91.1	
12 - 1	91.9	
13 - 1	91.8	
14 - 1	90	
16 - 1	90.8	
17 - 1	91	
18 - 1	92	
19 - 1	90	
20 - 1	91.2	
21 - 1	92	
23 - 1	91.2	
24 - 1	90.3	
25 - 1	91.6	
26 - 1	91.4	
27 - 1	90.3	
28 - 1	91.8	
30 - 1	89.9	

Westgard	تفسیر	C3	آزمایش	BS-480 اتوآنالایزر	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	91.1	SD	0.66	Exp. Date	1402/11/7	Lot. No	641-01-030/2
Next Month	Mean	91.1	SD	0.69	CV	0.76			

