

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	18.1	
3 - 1	18	1:2S
4 - 1	18.2	
5 - 1	17.6	
6 - 1	18.1	
7 - 1	17.7	
9 - 1	18	
10 - 1	18	
11 - 1	18.3	
12 - 1	17.4	
13 - 1	17.9	1:3S
14 - 1	18.1	
16 - 1	17.7	
17 - 1	18	
18 - 1	17.6	
19 - 1	17.4	
20 - 1	18	
21 - 1	17.4	
23 - 1	17.6	
24 - 1	18.2	
25 - 1	17.8	
26 - 1	17.9	
27 - 1	18.1	
28 - 1	17.6	
30 - 1	17.99	

Westgard	تفسیر	C4	آزمایش	BS-480 اتوآنالایزر	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	17.95	SD	0.28	Exp. Date	1402/9/14	Lot. No	641-01-030/2
Next Month	Mean	17.87	SD	0.27	CV	1.5			

