

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
12 - 1	60.3	
14 - 1	63	
15 - 1	62.7	
16 - 1	64.4	
17 - 1	60.3	
18 - 1	69.2	1:2S
24 - 1	56	
25 - 1	62	
26 - 1	55.2	
28 - 1	57.6	
29 - 1	51.6	
30 - 1	51.7	

Westgard	تفسیر	CA 19-9	آزمایش	2000 ایمولایت XPI-	دستگاه	11	ماه	1402	سال
Control	Seronorm	Mean	58.2	SD	3.97	Exp. Date	1402/11/15	Lot. No	
Next Month	Mean	59.5	SD	5.27	CV	8.86			

