

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	21.9	
3 - 1	22.1	
4 - 1	22.9	
5 - 1	22.4	
6 - 1	23.4	
7 - 1	23.9	
9 - 1	23.7	
10 - 1	22.1	
11 - 1	23.2	
12 - 1	23.4	
13 - 1	23.9	
14 - 1	23.4	
16 - 1	24.1	
17 - 1	24.1	
18 - 1	24.9	
19 - 1	24.1	
20 - 1	23.9	
21 - 1	22.8	
23 - 1	23.8	
24 - 1	23.9	
25 - 1	24.8	
26 - 1	22.1	
27 - 1	23.4	
28 - 1	23.9	
30 - 1	24.1	

Westgard	تفسیر	CEA	آزمایش	2000 ایمولایت XPI-	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	23.58	SD	1.18	Exp. Date	1403/8/30	Lot. No	36502028
Next Month	Mean	23.45	SD	0.83	CV	3.54			

