

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	127.27
SD	1.7
CV	1.33
Day-Run	Value
1 - 1	127
2 - 1	130
3 - 1	125
4 - 1	125
7 - 1	129
8 - 1	127
9 - 1	127
10 - 1	130
11 - 1	126
12 - 1	129
14 - 1	127
15 - 1	127
16 - 1	128
17 - 1	126
18 - 1	128
21 - 1	130
23 - 1	126
24 - 1	129
25 - 1	125
28 - 1	126
29 - 1	128
30 - 1	125

High - CV : 1.3	
Mean	217.23
SD	2.83
CV	1.3
Day-Run	Value
1 - 1	221
2 - 1	217
3 - 1	220
4 - 1	220
7 - 1	218
8 - 1	214
9 - 1	220
10 - 1	213
11 - 1	215
12 - 1	216
14 - 1	220
15 - 1	215
16 - 1	222
17 - 1	217
18 - 1	217
21 - 1	213
23 - 1	218
24 - 1	215
25 - 1	214
28 - 1	220
29 - 1	220
30 - 1	214

Westgard	تفسیر	CPK	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	127.08	SD	1.6	Exp. Date	1403/1/17	Lot. No	035

