

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.33	
Mean	127.08	Mean	216.6
SD	1.6	SD	2.87
CV	1.26	CV	1.33
Day-Run	Value	Day-Run	Value
2 - 1	127	2 - 1	212
3 - 1	126	3 - 1	212
4 - 1	129	4 - 1	218
5 - 1	126	5 - 1	215
6 - 1	126	6 - 1	219
7 - 1	127	7 - 1	213
9 - 1	130	9 - 1	220
10 - 1	125	10 - 1	218
11 - 1	130	11 - 1	215
12 - 1	125	12 - 1	215
13 - 1	128	13 - 1	217
14 - 1	128	14 - 1	216
16 - 1	128	16 - 1	219
17 - 1	126	17 - 1	225
18 - 1	130	18 - 1	219
19 - 1	126	19 - 1	217
20 - 1	127	20 - 1	215
21 - 1	127	21 - 1	215
23 - 1	126	23 - 1	216
24 - 1	125	24 - 1	214
25 - 1	129	25 - 1	217
26 - 1	125.1	26 - 1	214
27 - 1	128	27 - 1	217
28 - 1	126	28 - 1	217
30 - 1	127	30 - 1	220

Westgard	تفسیر	CPK	آزمایش	اتوآنالایزر BS-800	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	126.05	SD	1.77	Exp. Date	1403/1/17	Lot. No	035

