

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	46.6	
2 - 2	48.2	
3 - 1	47.1	
4 - 1	48.2	
8 - 1	46.3	
8 - 2	48.5	
9 - 1	46.5	
10 - 1	47.8	
12 - 1	48.6	
14 - 1	47.5	
15 - 1	48.3	
16 - 1	48.2	
17 - 1	47.3	
18 - 1	48	
21 - 1	46.4	
23 - 1	46.6	
24 - 1	47.1	
25 - 1	48.2	
28 - 1	46.6	
29 - 1	47	
30 - 1	47.8	

Westgard	تفسیر	CRP-quantitative	آزمایش	BS-480 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	47.61	SD	0.66	Exp. Date	1403/6/8	Lot. No	326-2-034/3
Next Month	Mean	47.47	SD	0.77	CV	1.63			

