

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.47	
Mean	9.47	Mean	13.45
SD	0.11	SD	0.2
CV	1.13	CV	1.47
Day-Run	Value	Day-Run	Value
2 - 1	9.5	2 - 1	13
3 - 1	9.6	3 - 1	13.6
4 - 1	9.5	4 - 1	13.5
5 - 1	9.5	5 - 1	13.7
6 - 1	9.6	6 - 1	13.7
7 - 1	9.5	7 - 1	13.6
9 - 1	9.6	9 - 1	13.3
10 - 1	9.4	10 - 1	13.4
11 - 1	9.4	11 - 1	13.2
12 - 1	9.2	12 - 1	13.2
13 - 1	9.5	13 - 1	13.3
14 - 1	9.3	14 - 1	13.4
16 - 1	9.6	16 - 1	13.2
17 - 1	9.4	17 - 1	13.6
18 - 1	9.5	18 - 1	13.4
19 - 1	9.5	19 - 1	13.8
20 - 1	9.3	20 - 1	13.5
21 - 1	9.7	21 - 1	13.5
23 - 1	9.5	23 - 1	13.5
24 - 1	9.4	24 - 1	13.6
26 - 1	9.6	26 - 1	13.3
27 - 1	9.4	27 - 1	13.5
28 - 1	9.3	28 - 1	13.3
30 - 1	9.4	30 - 1	13.7

Westgard	تفسیر	Calcium	آزمایش	BS-800 اتوآنالایزر	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	9.5	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

