

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	9.48
SD	0.14
CV	1.45
Day-Run	Value
1 - 1	9.5
2 - 1	9.3
3 - 1	9.3
4 - 1	9.5
7 - 1	9.3
8 - 1	9.6
9 - 1	9.4
10 - 1	9.4
11 - 1	9.5
12 - 1	9.5
14 - 1	9.7
15 - 1	9.3
16 - 1	9.5
17 - 1	9.3
18 - 1	9.6
21 - 1	9.4
23 - 1	9.6
24 - 1	9.7
25 - 1	9.6
28 - 1	9.4
29 - 1	9.4
30 - 1	9.7

High - CV : 1.7	
Mean	13.29
SD	0.23
CV	1.7
Day-Run	Value
1 - 1	13.4
2 - 1	13.5
3 - 1	13.2
4 - 1	13.2
7 - 1	13.6
8 - 1	13.5
9 - 1	13.2
10 - 1	13.2
11 - 1	13
12 - 1	13.5
14 - 1	13.2
15 - 1	13
16 - 1	13.4
17 - 1	13
18 - 1	13
21 - 1	13.7
23 - 1	13
24 - 1	13.4
25 - 1	13.5
28 - 1	13.1
29 - 1	13.6
30 - 1	13.2

Westgard	تفسیر	Calcium	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	9.47	SD	0.11	Exp. Date	1403/1/17	Lot. No	035

