

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	175.24
SD	2.24
CV	1.28
Day-Run	Value
2 - 1	170
3 - 1	176
4 - 1	176
5 - 1	175
6 - 1	174
7 - 1	173
9 - 1	177
10 - 1	175
11 - 1	176
12 - 1	176
13 - 1	173
14 - 1	174
16 - 1	172
17 - 1	176
18 - 1	177
19 - 1	180
20 - 1	173
21 - 1	176
23 - 1	177
24 - 1	173
25 - 1	173
26 - 1	178
27 - 1	176
28 - 1	178
30 - 1	177

High - CV : 0.93	
Mean	263.44
SD	2.45
CV	0.93
Day-Run	Value
2 - 1	267
3 - 1	262
4 - 1	264
5 - 1	264
6 - 1	260
7 - 1	264
9 - 1	260
10 - 1	262
11 - 1	264
12 - 1	266
13 - 1	267
14 - 1	261
16 - 1	262
17 - 1	265
18 - 1	266
19 - 1	267
20 - 1	263
21 - 1	265
23 - 1	263
24 - 1	261
25 - 1	261
26 - 1	260
27 - 1	265
28 - 1	267
30 - 1	260

Westgard	تفسیر	Cholesterol	آزمایش	اتوآنالایزر BS-800	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	174.95	SD	2.68	Exp. Date	1403/1/17	Lot. No	035

