

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	2.64
SD	0.04
CV	1.37
Day-Run	Value
1 - 1	2.64
2 - 1	2.65
3 - 1	2.6
4 - 1	2.66
7 - 1	2.67
8 - 1	2.63
9 - 1	2.7
10 - 1	2.62
11 - 1	2.69
12 - 1	2.67
14 - 1	2.63
15 - 1	2.6
16 - 1	2.65
17 - 1	2.6
18 - 1	2.66
21 - 1	2.69
23 - 1	2.7
24 - 1	2.6
25 - 1	2.61
28 - 1	2.61
29 - 1	2.6
30 - 1	2.69

High - CV : 1.43	
Mean	3.71
SD	0.05
CV	1.43
Day-Run	Value
1 - 1	3.75
2 - 1	3.76
3 - 1	3.72
4 - 1	3.7
7 - 1	3.7
8 - 1	3.6
9 - 1	3.7
10 - 1	3.6
11 - 1	3.8
12 - 1	3.65
14 - 1	3.7
15 - 1	3.7
16 - 1	3.7
17 - 1	3.8
18 - 1	3.7
21 - 1	3.7
23 - 1	3.7
24 - 1	3.8
25 - 1	3.72
28 - 1	3.7
29 - 1	3.7
30 - 1	3.75

Westgard	تفسیر	Creatinine	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	2.65	SD	0.04	Exp. Date	1403/1/17	Lot. No	035

