

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	130	1:2S
2 - 1	125	
3 - 1	126	
4 - 1	127	
7 - 1	128	
10 - 1	128	
11 - 1	127	
12 - 1	128	
14 - 1	129	
15 - 1	127	
16 - 1	126	
17 - 1	127	
18 - 1	128	
21 - 1	125	
23 - 1	128	
24 - 1	130	
25 - 1	127	
28 - 1	127	
29 - 1	125	
30 - 1	127	

Westgard	تفسیر	Copper (Cu)	آزمایش	BS-800 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	127.7	SD	1.5	Exp. Date	1402/11/30	Lot. No	hsn.35
Next Month	Mean	127.25	SD	1.45	CV	1.14			

