

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
3 - 1	25.3	
4 - 1	25.3	
7 - 1	24.5	
9 - 1	25.2	
10 - 1	25.6	1:2S
11 - 1	25	
12 - 1	25.1	
13 - 1	24.9	
14 - 1	24.6	
16 - 1	25.3	
17 - 1	25.1	
18 - 1	24.9	
19 - 1	25.1	
20 - 1	24.6	
21 - 1	24.8	
23 - 1	25	
24 - 1	24.6	
24 - 2	24.9	
25 - 1	25.2	
26 - 1	24.8	

Westgard	تفسیر	Cu,urine	آزمایش	Atomic Absorption Spectrometry	دستگاه	10	ماه	1402	سال
Control	آزمایشگاه	Mean	24.8	SD	0.35	Exp. Date	1402/11/30	Lot. No	
Next Month	Mean	24.99	SD	0.29	CV	1.15			

