

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	0.47
SD	0.03
CV	6.29
Day-Run	Value
2 - 1	0.45
3 - 1	0.43
4 - 1	0.48
5 - 1	0.46
6 - 1	0.5
7 - 1	0.49
9 - 1	0.43
10 - 1	0.48
11 - 1	0.45
12 - 1	0.48
13 - 1	0.51
14 - 1	0.5
14 - 1	0.5
16 - 1	0.42
17 - 1	0.48
18 - 1	0.52
19 - 1	0.49
20 - 1	0.46
21 - 1	0.52
23 - 1	0.42
24 - 1	0.44
25 - 1	0.48
26 - 1	0.49
27 - 1	0.45
28 - 1	0.48
30 - 1	0.5

High - CV : 3.21	
Mean	2.24
SD	0.07
CV	3.21
Day-Run	Value
2 - 1	2.3
3 - 1	2.23
4 - 1	2.25
5 - 1	2.3
6 - 1	2.11
7 - 1	2.3
9 - 1	2.13
10 - 1	2.31
11 - 1	2.17
12 - 1	2.27
13 - 1	2.23
14 - 1	2.1
16 - 1	2.31
17 - 1	2.1
18 - 1	2.18
19 - 1	2.3
20 - 1	2.3
21 - 1	2.2
23 - 1	2.3
24 - 1	2.3
25 - 1	2.2
26 - 1	2.3
27 - 1	2.21
28 - 1	2.25
30 - 1	2.3

Westgard	تفسیر	D.Bili	آزمایش	اتوآنالایزر BS-800	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	0.47	SD	0.04	Exp. Date	1403/1/17	Lot. No	035

