

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
15 - 1	22.58	1:2S
2 - 1	22.65	
2 - 2	22.35	
3 - 1	22.45	
4 - 1	22.89	
5 - 1	22.76	
6 - 1	22.74	
8 - 1	22.61	
9 - 1	22.35	
10 - 1	22.76	
11 - 1	22.75	
12 - 1	22.85	
13 - 1	22.61	
16 - 1	22.38	
17 - 1	22.45	
18 - 1	22.75	
19 - 1	22.64	
20 - 1	22.86	
23 - 1	22.53	
24 - 1	22.52	
25 - 1	22.68	
26 - 1	22.75	
27 - 1	22.35	
30 - 1	22.85	

Westgard	تفسیر	FSH	آزمایش	2000 ایمولایت XPI-	دستگاه	11	ماه	1401	سال
Control	biorex	Mean	22.63	SD	0.31	Exp. Date	1402/11/15	Lot. No	
Next Month	Mean	22.63	SD	0.17	CV	0.77			

